

The image features a large, abstract graphic on the left side, consisting of numerous thin, yellow, curved lines that sweep across a dark blue background. These lines originate from the left and curve towards the right, creating a sense of motion and depth. The lines are more densely packed in some areas, creating a mesh-like effect. The overall composition is modern and technical.


**bre**

**Performance testing of  
the Solar Star  
photovoltaic powered  
fan**

Prepared for: Ian Bell,  
SolaLighting Ltd

22 August 2008

Test report number 245-817



### Prepared by

---

Name Richard Kemp  
Position Senior Consultant  
Date 28 August 2008

Signature

---

### Approved on behalf of BRE

---

Name David Butler  
Position Manager, Building Diagnostics & HVAC Engineering  
Date 28 August 2008


Signature

---

BRE  
Garston  
WD25 9XX  
T + 44 (0) 1923 664000  
F + 44 (0) 1923 664010  
E [enquiries@bre.co.uk](mailto:enquiries@bre.co.uk)  
[www.bre.co.uk](http://www.bre.co.uk)


This report may only be distributed in its entirety and in accordance with the terms and conditions of the contract. Test results relate only to the items tested. BRE has no responsibility for the design, materials, workmanship or performance of the product or items tested. This report does not constitute an approval, certification or endorsement of the product tested.

This report is made on behalf of BRE. By receiving the report and action on it, the client – or any third party relying on it – accepts that no individual is personally liable in contract, tort or breach of statutory duty (including negligence).



## Contents


1	Introduction	4
2	Details of tests carried out	5
3	Test results	7



## 1 Introduction

Discussions between Ian Bell of SolaLighting and BRE representatives took place on the 9<sup>th</sup> May 2008 regarding Solalighting's requirements for testing and evaluation of the Solar Star photovoltaic-powered loft extract fan. Written acceptance of the work, as per BRE's proposal number 122460 dated 14<sup>th</sup> May 2008, was subsequently received by BRE. The work program comprised three tasks and this report details the results of the laboratory performance testing (Task 1 of the agreed works). Of interest to SolaLighting was the pressure and volume flow characteristics of the Solar Star under varying irradiance levels with a single and dual photovoltaic (PV) panel configuration.

A Solar Star fan, in the High Profile Roof Mount configuration, with integral PV and two additional PV panels (identical to the integrated panels) was provided by SolaLighting.



## 2 Details of tests carried out

The Solar Star supplied by SolaLighting was tested between the 27<sup>th</sup> July and 22<sup>nd</sup> August 2008.

The integral PV was disconnected and the additional PV panels provided were used in single and parallel electrical configurations for the purposes of the testing.


A schematic drawing of the purpose-built test rig is shown in **Figure 1** and selected photographs of the rig are given in Appendix A.

The rig was located internally in a laboratory with the exception of the solar components. The SolarStar was mounted at the end of the plenum chamber with the exhaust to the laboratory (draw through configuration). The PVs were mounted on a frame along with a Kipp and Zonen solarimeter in the same plane. A perpendicular rod was also mounted on this frame to check for incidence angle of the beam radiation. This frame could be moved to match the azimuth and altitude of the sun and throughout the tests the frame was always orientated normal to the sun's beam. The panels and the solarimeter were connected to the Solar Star and datalogger respectively by cables of 2.5mm<sup>2</sup> copper cores to ensure negligible voltage drop and resistive heating compared to the integrally mounted PV.

The tests were carried out in accordance with the recommendations of *BS 848 : Part 1 1997 ISO 5801 : 1997 Fans for general purposes Part 1. Performance testing using standardised airways* using the type A installation (inlet side test-chamber).

The plenum and supply ductwork was sealed, thus by installing air flow measuring devices in the makeup duct to the plenum, the extract rate of the SolarStar could be measured. The air flow rate supplied to the Solar Star was measured using a Furness FCO96H-2000L Laminar flow element or a 52.73mm or 70.84mm orifice diameter Poddymeter orifice plate, depending on the flowrates for the particular test condition, installed within a 104mm internal diameter duct. A Furness FCO12 differential pressure micromanometer and an Airflow Developments PVM100 micromanometer were used to measure pressure difference across the laminar flow element and the orifice plates respectively. The variables irradiance, plenum pressure, plenum temperature, room temperature and relative humidity were logged by a Grant 1000 Series Squirrel datalogger with the airflow measuring devices' differential pressure and atmospheric pressure recorded manually.


A variable speed auxiliary fan was installed at the measuring duct inlet. The speed was adjusted to create varying degrees of static pressure in the test plenum. Static pressure at this point was measured using a Furness micromanometer type FC012. Calibration certificates for the micromanometer instruments are shown in Appendix B to this report. A butterfly damper was installed at the inlet of the plenum to allow testing of low flow rates with higher differential pressures between the plenum and atmosphere when the auxiliary fan was switched off.



**Figure 1 Fan test rig (not drawn to scale)**

The temperature of the air at the Solar Star inlet was measured using a thermistor calibrated against a Hewlett Packard 2804A Quartz Thermometer. The relative humidity in the laboratory was measured by a Vaisala RH probe. A Calibration certificate for the quartz thermometer is reproduced in Appendix B to this report.


The tests were carried out at different levels of irradiance, namely 800 and 200W/m<sup>2</sup> (as measured on the beam-normal plane) as representative of a typical clear and overcast day during summer in the U.K. with an additional test series carried out at 500W/m<sup>2</sup> for two PVs in parallel. For each condition, the static pressure in the plenum was incrementally varied from the near-minimum achievable by closing the inlet damper, through to equalisation with the atmosphere by opening the inlet damper and increasing the auxiliary fan speed such that the SolarStar was operating against no pressure differential. For each plenum pressure, the SolarStar flowrate was measured via the pressure difference across the measuring device.






### 3 Test results

The fan tests were conducted over several days thus the plenum temperature, atmospheric pressure and relative humidity varied slightly from day to day but in all cases the values were measured and accounted for accordingly in the flow measurements.

The following charts show fan performance curves for the Solar Star with one PV panel and with two PV panels in parallel, for different levels of incident solar irradiance. A differential pressure of 0 Pa represents operation in free air with no inlet ducting.










## Appendix A Rig photographs




General layout of internal laboratory apparatus showing Solar Star test fan mounted on the plenum with orifice plates in the supply duct connected to the supply fan with variable speed drive





Above: plenum with Solar Star test fan




Frame with surface mounted PV panels, solarimeter and incidence angle measuring rod (shown indoors but in practice located externally in direct sunlight orientated normal to beam radiation)



## Appendix B Instrument calibration certificates

CERTIFICATE OF CALIBRATION	
DATE OF ISSUE 25 March 2008	CERTIFICATE NUMBER N1177276T
<b>ASAP Calibration Services Ltd</b> <b>Romsey Laboratory</b> UNIVERSAL HOUSE ROMSEY INDUSTRIAL ESTATE ROMSEY HAMPSHIRE SO51 0HR Telephone 01794 523935 Facsimile 01794 523910 Email info@asap-cal.co.uk Website www.asap-cal.co.uk	
	
UKAS CALIBRATION 0013	
PAGE 1 OF 5 PAGES	
APPROVED SIGNATORY [ ] C. C-Smith [x] D. Minton [ ] [ ]	
Customer :	BRE LTD
Address :	GARSTON GOODS INWARDS BUCKNALLS LANE, GARSTON WATFORD HERTS. WD25 9XX
Order No :	000226015
Apparatus Tested -	DIGITAL QUARTZ THERMOMETER & 2 RTD PROBES
Type No :	2804A, RTD
Serial No :	1744A00580, Z120A-01425, 1731A01110
Inventory No :	IN3895
Manufacturer :	HEWLETT PACKARD
Range/Scale :	CALIBRATED RANGE: 0 TO 85 °C, 0.01 °C RESOLUTION
Test Conditions -	
Date Instrument Received :	13 March 2008
Date Calibration Completed :	25 March 2008
Ambient Temperature :	20 ± 2 °C
Reference No :	1177276
Certified by	
<small>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</small>	
22-005-079/009/FEB2007	



## CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0013

CERTIFICATE NUMBER

N1177276T

PAGE 2 OF 5 PAGES

CALIBRATED "AS FOUND" {i.e. NO ADJUSTMENTS MADE}

CHANNEL: T1 THUMB-WHEEL SETTING: 540 (AS FOUND)

SERIAL NO: INDICATOR:1744A00580 PROBE:Z120A-01425

THE INSTRUMENT WAS CALIBRATED BY IMMERSING THE PROBE IN CLOSELY CONTROLLED TEMPERATURE REFERENCE ENVIRONMENTS TOGETHER WITH TWO REFERENCE STANDARD INSTRUMENTS HAVING KNOWN AND TRACEABLE VALUES OF UNCERTAINTY. THE INSTRUMENT WAS CALIBRATED BY USING THE COMPARISON METHOD.

THE FOLLOWING RESULTS ARE DERIVED FROM THE MEAN VALUES OF A NUMBER OF OBSERVATIONS.

THE PROBE WAS IMMERSED TO A MINIMUM DEPTH OF 200mm.

TEST TEMPERATURE °C	INSTRUMENT READING °C
0.003	0.02
10.030	10.05
20.031	20.06
40.004	40.04
60.039	60.09
85.169	85.26
0.002	0.03


THE UNCERTAINTY ASSOCIATED WITH ABOVE MEASUREMENTS IS  
 $\pm [0.01^{\circ}\text{C} + \text{INSTRUMENT RESOLUTION}]$ .

THE ABOVE UNCERTAINTY REFERS TO THE MEASUREMENT AND IS NOT INTENDED TO INDICATE THE SPECIFICATION, OR REPEATABILITY OF THE INSTRUMENT.

THE TEMPERATURE SCALE IN USE IS THE INTERNATIONAL TEMPERATURE SCALE OF 1990.  
ITS - 90.

Test Engineer

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2.00$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.



# CERTIFICATE OF CALIBRATION

Technical & Calibration Services

TITLE: FURNESS MICROMANOMETER

CERTIFICATE N° : 4197

INSTRUMENT TYPE: FCO12

T.S. JOB N° : 51003

SERIAL N° : 9002509

BRE N° : IN 3986

CALIBRATION PROCEDURE: SP5/08/R

DATE OF TEST: 25 SEP 07

NEXT CALIBRATION DUE: 25 SEP 08

CALIBRATION EQUIPMENT: FURNESS FCO510 MICROMANOMETER, SERIAL N° 0407141, BRE N° IN3307, UKAS CALIBRATION CERTIFICATE N° 03972, DATED 28 AUG 07.

FCO510 APPLIED TRUE PRESSURE Pa	MICROMANOMETER INDICATED PRESSURE, 10%Δp range				
	UP Pa	DOWN Pa	UP Pa	DOWN Pa	MEAN Pa
0	±0.01	±0.01	±0.01	±0.01	±0.02
5	4.93	4.98	4.94	4.94	4.87
10	9.90	9.90	9.88	9.89	9.75
15	14.83	14.85	14.81	14.84	14.61
20	19.77		19.75		19.49
Pa	100%Δp range				
	Pa	Pa	Pa	Pa	Pa
0	0	0	0	0	0
20	19.6	19.8	19.6	19.8	20.3
30	29.5	29.7	29.4	29.6	30.4
40	39.3	39.6	39.3	39.5	40.4
50	49.2	49.5	49.2	49.4	50.7
100	98.5	98.7	98.4	98.6	100.8
150	147.7	148.0	147.6	147.7	150.4
200	197.0		196.8		199.3

The expanded uncertainty of the micromanometer indicated pressures, after correction, is stated in two parts:  
Relative uncertainty: ±0.13% for the 10%Δp range and ±0.14% for the 100%Δp range,  
Absolute uncertainty: ±0.6Pa for the 10%Δp range and ±1.2Pa for the 100%Δp range.

The reported two-part expanded uncertainty is in each case based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. For each stated result the user may, if required, combine the uncertainties shown by quadrature summation in either relative or absolute units as appropriate.

Calibrated By: S. C. Wright


Test Area Temperature: 20°C ±2°C

Signature:

Date:

Test report number 245-817  
Commercial in confidence

© Building Research Establishment Ltd 2008  
Page 13 of 15



# CERTIFICATE OF CALIBRATION

Technical & Calibration Services

TITLE: AIRFLOW MICROMANOMETER

CERTIFICATE N° : 4198

INSTRUMENT TYPE: PVM100

T.S. JOB N° : 51003

SERIAL N° : 079716

BRE N° : IN 3987

CALIBRATION PROCEDURE: SP5/08/R

DATE OF TEST: 26 SEP 07

NEXT CALIBRATION DUE: 26 SEP 08

CALIBRATION EQUIPMENT: FURNESS FCO510 MICROMANOMETER, SERIAL N° 0407141, BRE N° IN3307, UKAS CALIBRATION CERTIFICATE N° 03972, DATED 28 AUG 07.  
BUDENBERG AIR DEAD WEIGHT TESTER TYPE 451, SERIAL N°s 126E & 3145, BRE N°s IN1491, 2 & 3, UKAS CALIBRATION CERTIFICATE N° 09272, DATED 13 SEP 05.

FCO510 & D.W.T. APPLIED PRESSURE Pa	MICROMANOMETER INDICATED PRESSURE			
	UP Pa	DOWN Pa	UP Pa	DOWN Pa
0	0	0	0	1
50.0	50	50	50	50
100.0	99	100	100	100
200.0	199	199	199	200
400.0	399	399	399	400
600.0	599	600	599	600
800.0	800	801	800	801
1000.0	1001	1001	1001	1001
1500.8	1504	1505	1505	1506
2001.1	2005	2006	2006	2006
2501.3	2505	2506	2506	2506
3001.6	3007	3007	3008	3008
3501.9	3510		3511	

The above deadweight tester applied pressure values (1500 to 3500Pa) have been corrected for local gravity.

The expanded uncertainty of the micromanometer indicated pressures, after correction, is stated in two parts:

Relative uncertainty:  $\pm 0.13\% \leq 1000\text{Pa} < \pm 0.05\%$ ,

Absolute uncertainty:  $\pm 3.0\text{ Pa}$ .

The reported two-part expanded uncertainty is in each case based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. For each stated result the user may, if required, combine the uncertainties shown by quadrature summation in either relative or absolute units as appropriate.


Calibrated By: S. C. Wright

Test Area Temperature:  $20^\circ\text{C} \pm 2^\circ\text{C}$

Signature:  
Test report number 245-817  
Commercial in confidence

Date:

© Building Research Establishment Ltd 2008  
Page 14 of 15



=====REPORT ENDS=====